



NJ-1608

- Great consulting comfort, value for your money. Need based unit, with slide tray for trial lens set, side table for writing or lens meter, takes two main instruments, place for chart projector, overhead light, comfortable polyurethane seat, holder for hand instruments, control are for foot or side of the chair.
- Option of manual or motorized recline.
- Option of 110v or 220v.
- Option of Surgeons Stool.
- Option of Phoropter arm.

Further measurements	
Height of projector column	130-140 cm
Height of seat on chair (lowest setting)	56 cm
Height of seat on chair (highest setting)	76 cm

Weight	
Maximum weight of equipment without additional apparatus	approx. 150 kg

Electrical date for refraction unit	
Mains voltage	230 V AC
Mains frequency	50 / 60 Hz
Connection power	400 VA

REF | TE/CAT/NJ1608/31/06052024

TARUN ENTERPRISES

H.O.: 8/8, Strachy Road, Prayagraj – 211001, U.P., INDIA
 Phone: +91 - 8176080204, Email: info@optitecheyecare.com

B.O.: A96/1, Second Floor, Phase II, Mayapuri Industrial Area, New Delhi – 110064
 Phone: 011-47070362, Email: delhi@optitecheyecare.com